



Remarks :

1. After the Loading Screw (item# 12) is set to the desired level, apply a little silicone to the Loading Screw from underneath so that its level will not change after a period of time.
2. Two positioning bolts should be on the center of two top plate holes and below top level 3mm after two nuts are fasterned. The bolt can not tarin top plate.
3. Recmmended faserern torque for load cell is 90Nm~110Nm

	APPROVED		DATE	
V ₂	APPROVED Gary.G	DATE	09-08-26	
	CHECKED Eric.Q	DATE	09-08-26	
	DRAWN Fred.Z	DATE	10-09-09	

14	Shield			1	
13	Plate 4			1	
12	Loading Screw			1	
11	Cylinder Ring			1	
10	Loading Pin			1	
9	Cap			1	
8	Nut	M12		2	
7	Bolt 3	M12X100		2	
6	Plate 2			1	
5	Bolt 2	M6X10		2	
4	Bolt 1	M12X55		2	
3	Washer	φ12		2	
2	Load Cell			1	563YH
1	Plate 1			1	
ITEM	DESCRIPTION	MATERIAL	SPECIFICATION	QTY	REMARKS

ANYLOAD

TITLE:
563YHM5-(0.5-1)K1b
General Arrangement

DRAWING NO. :
 PROJECTION: SHEET 1 OF 1

IMPORTANT NOTICE

This document consisting confidential proprietary information belongs to ANYLOAD who retains all copy and other rights. This document is submitted under a confidential relationship for specific purpose, and the receipt agrees by accepting this document not to supply or disclose any information regarding it to any unauthorized person.